Data File C:\HPCHEM\1\DATA\N9709231\037B3701.D Sample Name: 000945162880

Injection Date : 9/23/97 8:35:48 PM Seq. Line : 37
Sample Name : 000945162880 Vial : 37
Acq. Operator : martinj Inj : 1
Inj Volume : 1 µl

Acq. Method : C:\HPCHEM\1\METHODS\H2O.M
Last changed : 9/23/97 11:43:59 AM by martinj
Analysis Method : C:\HPCHEM\1\METHODS\NIC.M
Last changed : 10/3/97 10:41:00 AM
(modified after loading)
Water Analysis - Routine Nicotine and Water Analysis



Internal Standard Report (Sample Amount is 0!)

Sorted By : Signal
Calib. Data Modified : 10/3/97 10:35:20 AM
Multiplier : 1.0000
Dilution : 1.0000
Uncalibrated Peaks : not reported

Sample ISTD Information:

ISTD #	ISTD Amount [mg/cigt]	Name
1	1.00000	Carvone

Signal 1: FID2 B,
Results obtained with enhanced integrator!

RetTime [min]	Type	Area counts*s	Amt/Area ratio	Amount [mg/cigt]	Grp	Name
2.822	VV T+I	2.74292e4	1.00000	1.00000		Carvone
3.654		-	-	-		Nic

Totals without ISTD(s) : 0.00000

2 Warnings or Errors :

Warning : Time reference compound(s) not found
Warning : Sample amount is zero. Absolute amounts calculated

PM3000376369

Source: https://www.industrydocuments.ucsf.edu/docs/rlwl0001